

Quality Assurance/Quality Control and Data Process



Asia Pacific Mercury
Monitoring Network



Center for Environmental
Monitoring and Technology
National Central University

Quality Control

Quality control (QC) tests demonstrating the performance of the modified method, including the following:

- Calibration
- Initial precision and recovery
- Analysis of blanks
- Matrix spike/matrix spike duplicate analysis
- Ongoing precision and recovery
- Quality control sample
- Method detection limit


Analytical Batch

- a. Three blanks
- b. Five calibration standards
- c. Ongoing precision and recovery (OPR)
- d. Quality control sample
- e. Method blank
- f. **Seven samples**
- g. Method blank
- h. **Three samples**
- i. Matrix spike/Matrix spike duplicate
- j. **Four samples**
- k. Method blank
- l. **Six samples**
- m. Matrix spike/Matrix spike duplicate
- n. Ongoing precision and recovery

Analytical Batch


- a. Three blanks
- b. Five calibration standards
- c. Ongoing precision and recovery (OPR)
- d. Quality control sample
- .
- .
- .
- n. Ongoing precision and recovery

Unverified



Verified

12 hours has elapsed



Contamination Control

The philosophy behind contamination control is to ensure that any object or substance that contacts the sample is **metal free and free from any material that may contain mercury**

Contamination Control

- Avoiding contamination
- Use a clean environment
- Minimize exposure
- Clean work surfaces
- Wear gloves
- Use metal-free Apparatus
- Avoid sources of contamination

System calibration

1. Bubbler system calibration

- Establish the operating conditions necessary to purge Hg from the bubbler and to desorb Hg from the traps in a sharp peak => The calibration must contain **a minimum of 5 non-zero points** and the results of analysis of **three bubbler blanks**.
- Analyze the standards beginning with the lowest concentration and proceeding to the highest.

System calibration

1. Bubbler system calibration

- Prepare and analyze a minimum of 3 bubbler blanks. If the result for any method blank is found to **contain > 0.50 ng/L (50 pg) Hg**, the system is out of control.
- Calculate the calibration factor (CF_x) for Hg in each of the five standards using the mean bubbler-blanksubtracted peak height or area.

System calibration

$$CF_x = \frac{(A_x) - (\bar{A}_{BB})}{(C_x)}$$

Where:

A_x = peak height or area for Hg in standard

$\bar{A}_{BB}$ = mean peak height or area for Hg in bubbler blank

C_x = mass in standard analyzed (ng)

Calculate the mean calibration factor (CFm), SD calibration factor, RSD of the calibration factor,

where $RSD = 100 \times SD/CFm$.

If **RSD < 15%**, calculate the recovery for the **lowest standard** using CFm. If the recovery of the lowest standard is in the range of **75-125%**, the calibration is **acceptable**.

System calibration

2. Flow-injection system calibration

- The criteria are similar with the bubbler system calibration. Only difference in repairing the solution.

$$CF_x = \frac{(A_x) - (\bar{A}_{SB})}{(C_x)}$$

Where:

A_x = peak height or area for Hg in standard

$\bar{A}_{SB}$ = mean peak height or area for Hg in calibration blanks

C_x = concentration of standard analyzed (ng/L)

Calibration curve

- Calibration to a range other than 0.5 to 100 ng/L
- There must be a minimum of five non-zero calibration points
- The difference between successive calibration points must be no greater than a factor of 10 and no less than a factor of 2.
- $RSD < 15\%$. The calibration factor for any calibration point at a concentration greater than 100 ng/L must be within $\pm 15\%$ of the average calibration factor for the points at or below 100 ng/L.
- The calibration factor for any point < 0.5 ng/L must be within 25% of the average calibration factor for all points.
- Our calibration curve : 0.5, 1, 5, 10, 25, 50 ng/L

Analysis of blanks

- Bubbler blanks

*demonstrate that **bubbler systems** are free from contamination at levels that could affect data quality*

- System blanks

*demonstrate that **flow injection systems** are free from contamination at levels that could affect data quality*

Analysis of blanks

- Reagent blanks

*demonstrate that the **reagents** used to prepare samples for Hg analyses are **free from contamination***

- Method blanks

*demonstrate that the **analytical system** is free from contamination that could otherwise compromise sample results*

Analysis of blanks

- Field blanks

demonstrate that samples have not been contaminated by the **sample collection and transport activities**

- Equipment blanks

demonstrate that the **sampling equipment** is free from contamination

- Bottle blanks

verify the effectiveness of the **cleaning procedures**

Criteria for Performance

Quality Control Acceptance Criteria for Performance Tests in EPA Method 1631

Table 2

Acceptance Criteria	Section	Limit (%)
Initial Precision and Recovery (IPR)	9.2.2	
Precision (RSD)	9.2.2.3	21
Recovery (X)	9.2.2.3	79-121
Ongoing Precision and Recovery (OPR)	9.5.2	77-123
Matrix Spike/Matrix Spike Duplicate (MS/MSD)	9.3	
Recovery	9.3.4	71-125
Relative Percent Difference (RPD)	9.3.5	24

Initial precision and recovery

*To establish the **ability to generate acceptable precision and recovery**, the laboratory shall perform the following operations*

1. Analyze four replicates of the IPR solution (5 ng/L)
2. Using the results of the set of four analyses, compute the average percent recovery ($\bar{X}$), and the standard deviation of the percent recovery (s) for Hg
3. Compare s and $\bar{X}$ with the corresponding limits for initial precision and recovery. (**79 – 121% (Accuracy) and 21% RSD (Precision)**)

Ongoing precision and recovery

To demonstrate that the analytical system is within the performance criteria of this method and that acceptable precision and recovery is being maintained within each analytical batch

Ongoing precision and recovery

Analyze the OPR solution (5 ng/L) prior to the analysis of each analytical. An OPR also must be analyzed **at the end of an analytical sequence or at the end of each 12-hour shift**

Ongoing precision and recovery

If the recovery is in the range specified, the analytical system is in control and **analysis of samples and blanks may proceed**. If, however, the concentration is not in the specified range, the analytical process is not in control.

Initial Precision and Recovery (IPR) and Ongoing Precision and Recovery (OPR) solutions

It can be the working Hg standard. Prepare IPR and OPR solutions at a concentration of 5 ng/L Hg in reagent water. IPR/OPR solutions are prepared using the same amounts of reagents used for preparation of the calibration standards.

Matrix spike/matrix spike duplicate analysis

To assess the performance of the Method on a given sample matrix, the laboratory must spike, in duplicate, a minimum of 10% (1 sample in 10) from a given sampling site or, if for compliance monitoring, from a given discharge.

The concentration of the spike in the sample shall be determined as 1–5 times the background concentration of the sample or at 1-5 times the ML in Table

Lowest Ambient Water Quality Criterion for Mercury and the Method Detection Limit and Minimum Level of Quantitation for EPA Method 1631

Metal	Lowest Ambient Water Quality Criterion ⁽¹⁾	Method Detection Limit (MDL) and Minimum Level (ML)	
		MDL ⁽²⁾	ML ⁽³⁾
Mercury (Hg)	1.3 ng/L	0.2 ng/L	0.5 ng/L

1. Lowest water quality criterion for the Great Lakes System (Table 4, 40 CFR 132.6).
The lowest Nationwide criterion is 12 ng/L (40 CFR 131.36).
2. Method detection limit (40 CFR 136, Appendix B)
3. Minimum level of quantitation (see Glossary)

Matrix spike/matrix spike duplicate analysis

Determine the background concentration (B). Spike two additional sample aliquots with identical amounts of the spiking solution and analyze these aliquots to determine the concentration after spiking (A)

$$\% R = 100 \frac{(A-B)}{T}$$

where:

A = Measured concentration of analyte after spiking
B = Measured concentration of analyte before spiking
T = True concentration of the spike

Matrix spike/matrix spike duplicate analysis

Compare percent recovery (R) with the QC acceptance criteria. If the results of both the spike and the OPR test **fall outside the acceptance criteria**, the analytical system **is judged to be not in control**, and the results may not be reported or used for permitting or regulatory compliance purposes


Matrix Spike/Matrix Spike Duplicate (MS/MSD)	9.3	
Recovery	9.3.4	71-125
Relative Percent Difference (RPD)	9.3.5	24

Quality control sample

The laboratory must obtain a QCS from a **source different** from the Hg used to produce the standards used routinely.

The QCS should be analyzed as an **independent check** of system performance.

Quality control sample



National Research Council Canada

[Home](#) > [Programs and services](#) > [Technical and advisory services](#) > [Certified reference materials \(CRMs\)](#)

> [List of products / Registry of Documentation](#) > ORMS-5: Elevated Mercury in River Water

ORMS-5: Elevated Mercury in River Water

Product documentation

- [ORMS-5](#) (PDF, 150 KB)

ORMS-5 is a river water spiked with inorganic mercury. The material is packaged in 50 ml glass ampoules stabilized with 0.5% BrCl.

Table 1: Certified quantity value

Element	pg/g
Hg	26.2 ± 1.3

The certified value was derived gravimetrically and corroborated with two methods of analysis: cold vapour isotope dilution inductively coupled plasma mass spectrometry (ID-ICP-MS) and flow injection cold vapour atomic absorption spectrometry (CV-AAS). The certified value is the unweighted mean of the results $\pm$ the expanded ($k=2$) uncertainty (U_{CRM}). The results generated by each of these methods were judged to be independent.

The expanded uncertainty in the certified value is equal to $U_{CRM} = k u_C$, where u_C is the combined standard uncertainty calculated according to the JCGM Guide and k is the coverage factor. It is intended that U_{CRM} encompasses the uncertainty of every factor that reasonably contributes to the total uncertainty of the measurand. A coverage factor $k=2$ is used to give an uncertainty interval that contains roughly 95% of the underlying distribution. The value of u_C is determined from the combined uncertainties of the methods (u_{char}) as well as uncertainties associated with homogeneity (u_{hom}).

Intended use

This CRM is intended for the calibration of instruments and evaluation of methods for the determination of mercury.

Date of issue: September 2011

Date of expiry: September 2021

Revised: March 2016 (editorial update)

► [Report a problem or mistake on this page](#)

Date modified: 2016-12-13

Method Detection Limit

To establish the ability to detect Hg, the laboratory shall achieve an MDL that is **less than or equal to the MDL (0.16 ng/L @ NCU)**

Lowest Ambient Water Quality Criterion for Mercury and the Method Detection Limit and Minimum Level of Quantitation for EPA Method 1631

Metal	Lowest Ambient Water Quality Criterion ⁽¹⁾	Method Detection Limit (MDL) and Minimum Level (ML)	
		MDL ⁽²⁾	ML ⁽³⁾
Mercury (Hg)	1.3 ng/L	0.2 ng/L	0.5 ng/L

1. Lowest water quality criterion for the Great Lakes System (Table 4, 40 CFR 132.6).
The lowest Nationwide criterion is 12 ng/L (40 CFR 131.36).
2. Method detection limit (40 CFR 136, Appendix B)
3. Minimum level of quantitation (see Glossary)

Data Analysis and Calculations

Calculations for the bubbler system

$$[Hg] \text{ (ng/L)} = \frac{A_s - \bar{A}_{BB}}{CF_m \times V}$$

where:

A_s = peak height (or area) for Hg in sample

$\bar{A}_{BB}$ = peak height (or area) for Hg in bubbler blank

CF_m = mean calibration factor (Section 10.2.2.5)

V = Volume of sample (L)

We can use this formula to calculate for concentration of Hg in method blanks, field blanks, and reagent blanks

Data Analysis and Calculations

Calculations for the flow-injection system

$$[\text{Hg}] \text{ (ng/L)} = \frac{(A_s - \bar{A}_{SB})}{CF_m} \times \frac{V_{std}}{V_{sample}}$$

where:

A_s = *peak height (or area) for Hg in sample*

$\bar{A}_{SB}$ = *mean peak height (or area) for Hg in system blanks*

CF_m = *mean calibration factor (Section 10.3.2.6)*

V_{std} = *volume (mL) used for standards – volume (mL) reagent used in standards*

V_{sample} = *volume (mL) of sample – volume (mL) reagent used in sample*